



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Transcat – Ireland
Unit 15, Ballytrasna Business Park,
Little Island, Cork, Ireland T45 DA34

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

A handwritten signature in black ink, appearing to be 'Jason Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 07 September 2025

Certificate Number: AC-2489.29



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Transcat – Ireland

Unit 15, Ballytrasna Business Park,
Little Island, Cork, Ireland T45 DA34
Kevin Nuss +353 021 4506572

CALIBRATION

Valid to: **September 7, 2025**

Certificate Number: **AC-2489.29**

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source ¹	Up to 120 μ A		Fluke 5560A Multiproduct Calibrator
	(3 to 45) Hz	0.02 % of reading + 7.8 nA	
	45 Hz to 1 kHz	0.02 % of reading + 7.8 nA	
	(1 to 5) kHz	0.02 % of reading + 7.8 nA	
	(5 to 10) kHz	0.12 % of reading + 32 nA	
	(10 to 30) kHz	0.39 % of reading + 0.78 μ A	
	(0.12 to 1.2) mA		
	(3 to 45) Hz	0.02 % of reading + 78 nA	
	45 Hz to 1 kHz	0.02 % of reading + 78 nA	
	(1 to 5) kHz	0.02 % of reading + 78 nA	
	(5 to 10) kHz	0.12 % of reading + 78 nA	
	(10 to 30) kHz	0.39 % of reading + 3.9 μ A	
	(1.2 to 12) mA		
	(3 to 45) Hz	0.02 % of reading + 0.78 μ A	
	45 Hz to 1 kHz	0.02 % of reading + 0.78 μ A	
	(1 to 5) kHz	0.02 % of reading + 0.78 μ A	
	(5 to 10) kHz	0.12 % of reading + 0.78 μ A	
	(10 to 30) kHz	0.39 % of reading + 7.8 μ A	
	(12 to 120) mA		
	(3 to 45) Hz	0.02 % of reading + 7.8 μ A	
	45 Hz to 1 kHz	0.012 % of reading + 3.9 μ A	
(1 to 5) kHz	0.02 % of reading + 6.3 μ A		
(5 to 10) kHz	0.12 % of reading + 7.8 μ A		
(10 to 30) kHz	0.39 % of reading + 78 μ A		

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source ¹	(0.12 to 1.2) A (3 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz (1.2 to 3.1) A (3 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (3.1 to 12) A (3 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (12 to 30,2) A (3 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz	0.02 % of reading + 78 μ A 0.02 % of reading + 39 μ A 0.02 % of reading + 63 μ A 0.12 % of reading + 0.24 mA 0.39 % of reading + 0.24 mA 0.03 % of reading + 0.39 mA 0.024 % of reading + 0.24 mA 0.03 % of reading + 0.24 mA 0.2 % of reading + 0.39 mA 0.03 % of reading + 0.78 mA 0.024 % of reading + 0.39 mA 0.03 % of reading + 0.63 mA 0.2 % of reading + 0.78 mA 0.078 % of reading + 7.8 mA 0.055 % of reading + 6.3 mA 0.39 % of reading + 6.3 mA	Fluke 5560A Multiproduct Calibrator
AC Current – Measure ¹	(0.2 to 20) μ A 1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (20 to 200) μ A 1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (0.2 to 2) mA 1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (2 to 20) mA 1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz	0.2 % of reading + 2.5 nA 0.2 % of reading + 2.5 nA 0.2 % of reading + 2.5 nA 0.028 % of reading + 5 nA 0.053 % of reading + 5 nA 0.074 % of reading + 5 nA 0.041 % of reading + 10 nA 0.028 % of reading + 50 nA 0.053 % of reading + 50 nA 0.074 % of reading + 50 nA 0.41 % of reading + 0.1 μ A 0.028 % of reading + 0.5 μ A 0.053 % of reading + 0.5 μ A 0.074 % of reading + 0.5 μ A 0.41 % of reading + 1 μ A	Fluke 8588A 8.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure ¹	(20 to 200) mA		Fluke 8588A 8.5 Digit Multimeter
	1 Hz to 2 kHz	0.028 % of reading + 5 μA	
	(2 to 10) kHz	0.052 % of reading + 5 μA	
	(10 to 30) kHz	0.074 % of reading + 5 μA	
	(0.2 to 2) A		
	1 Hz to 2 kHz	0.03 % of reading + 0.1 mA	
	(2 to 10) kHz	0.056 % of reading + 0.1 mA	
	(10 to 30) kHz	0.08 % of reading + 0.1 mA	
	(2 to 20) A		
	10 Hz to 2 kHz	0.084 % of reading + 0.5 mA	
(2 to 10) kHz	0.086 % of reading + 0.5 mA		
AC Clamp-on Ammeter (Toroidal Type) Transformer Type Sensor ¹	(20 to 60) A		Fluke 5560A Multiproduct Calibrator, Fluke 55XXA/COIL 50 50-turn Coil
	(45 to 440) Hz	0.5 % of reading + 2 mA	
	(60 to 155) A		
	(45 to 440) Hz	0.5 % of reading + 12 mA	
	(155 to 600) A		
(45 to 440) Hz	0.5 % of reading + 20 mA		
(600 to 1 500) A			
(45 to 440) Hz	0.51 % of reading + 0.31 A		
AC Clamp-on Ammeter (Non-Toroidal Type) Hall Effect Sensor ¹	(20 to 60) A		Fluke 5560A Multiproduct Calibrator, Fluke 55XXA/COIL 50 50-turn Coil
	(45 to 440) Hz	0.53 % of reading + 2 mA	
	(60 to 155) A		
	(45 to 440) Hz	0.53 % of reading + 12 mA	
	(155 to 600) A		
(45 to 440) Hz	0.53 % of reading + 20 mA		
(600 to 1 500) A			
(45 to 440) Hz	0.53 % of reading + 0.31 A		
AC Voltage – Source ¹	Up to 12 mV		Fluke 5560A Multiproduct Calibrator
	(3 to 5) Hz	0.2 % of reading + 5.5 μV	
	(5 to 10) Hz	0.068 % of reading + 5.5 μV	
	10 Hz to 20 kHz	0.012 % of reading + 4.7 μV	
	(20 to 50) kHz	0.03 % of reading + 4.7 μV	
	(50 to 100) kHz	0.12 % of reading + 12 μV	
	(100 to 300) kHz	0.63 % of reading + 24 μV	
	(300 to 500) kHz	0.63 % of reading + 24 μV	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source ¹	(12 to 120) mV		Fluke 5560A Multiproduct Calibrator
	(3 to 5) Hz	0.2 % of reading + 5.5 μV	
	(5 to 10) Hz	0.068 % of reading + 5.5 μV	
	10 Hz to 20 kHz	0.011 % of reading + 4.7 μV	
	(20 to 50) kHz	0.028 % of reading + 6.3 μV	
	(50 to 100) kHz	0.063 % of reading + 16 μV	
	(100 to 300) kHz	0.16 % of reading + 24 μV	
	(300 to 500) kHz	0.16 % of reading + 24 μV	
	(0.12 to 1.2) V		
	(3 to 5) Hz	0.2 % of reading + 59 μV	
	(5 to 10) Hz	0.068 % of reading + 55 μV	
	(10 to 40) Hz	0.011 % of reading + 47 μV	
	40 Hz to 20 kHz	0.011 % of reading + 6.3 μV	
	(20 to 50) kHz	0.024 % of reading + 11 μV	
	(50 to 100) kHz	0.055 % of reading + 32 μV	
	(100 to 300) kHz	0.15 % of reading + 63 μV	
	(300 to 500) kHz	0.15 % of reading + 63 μV	
	(1.2 to 12) V		
	(3 to 5) Hz	0.2 % of reading + 0.59 mV	
	(5 to 10) Hz	0.068 % of reading + 0.59 mV	
	(10 to 40) Hz	0.011 % of reading + 0.28 mV	
	40 Hz to 20 kHz	0.011 % of reading + 39 μV	
	(20 to 50) kHz	0.024 % of reading + 39 μV	
	(50 to 100) kHz	0.055 % of reading + 98 μV	
(100 to 300) kHz	0.16 % of reading + 0.47 mV		
(300 to 500) kHz	0.16 % of reading + 0.47 mV		
(12 to 70) V			
(3 to 5) Hz	0.2 % of reading + 5.9 mV		
(5 to 10) Hz	0.068 % of reading + 5.9 mV		
(10 to 40) Hz	0.011 % of reading + 2.8 mV		
40 Hz to 20 kHz	0.011 % of reading + 0.39 nV		
(20 to 50) kHz	0.024 % of reading + 0.39 mV		
(50 to 100) kHz	0.055 % of reading + 0.97 mV		
(100 to 300) kHz	0.14 % of reading + 16 mV		
(70 to 120) V			
(3 to 5) Hz	0.2 % of reading + 5.9 mV		
(5 to 10) Hz	0.068 % of reading + 5.9 mV		
(10 to 40) Hz	0.011 % of reading + 2.8 mV		
40 Hz to 20 kHz	0.011 % of reading + 0.39 mV		
(20 to 50) kHz	0.024 % of reading + 0.39 mV		
(50 to 100) kHz	0.055 % of reading + 0.97 mV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source ¹	(120 to 330) V (3 to 5) Hz (5 to 10) Hz 10 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (330 to 1 020) V (3 to 5) Hz (5 to 10) Hz 10 Hz to 10 kHz	0.2 % of reading + 59 mV 0.068 % of reading + 59 mV 0.011 % of reading + 6.3 mV 0.024 % of reading + 6.3 mV 0.12 % of reading + 9.7 mV 0.2 % of reading + 59 mV 0.068 % of reading + 59 mV 0.011 % of reading + 63 mV	Fluke 5560A Multiproduct Calibrator
AC Voltage – Measure ¹	(0.1 to 10) mV 1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (10 to 100) mV 1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz (0.1 to 1) V 1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	0.029 % of reading + 1.1 μV 0.037 % of reading + 1.1 μV 0.038 % of reading + 1.1 μV 0.3 % of reading + 0.78 μV 1 % of reading + 3.9 μV 2 % of reading + 3.9 μV 0.008 9 % of reading + 0.5 μV 0.013 % of reading + 0.5 μV 0.023 % of reading + 1 μV 0.053 % of reading + 5 μV 0.21 % of reading + 31 μV 1 % of reading + 0.1 mV 1.5 % of reading + 0.5 mV 4.1 % of reading + 1 mV 8.4 % of reading + 1 mV 16 % of reading + 1 mV 0.007 7% of reading + 5 μV 0.012 % of reading + 5 μV 0.023 % of reading + 10 μV 0.053 % of reading + 50 μV 0.21 % of reading + 0.31 mV 1 % of reading + 1 mV 1.5 % of reading + 5 mV 4 % of reading + 10 mV 8.2 % of reading + 10 mV 15 % of reading + 10 mV	Fluke 8588A 8.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure ¹	(1 to 10) V		Fluke 8588A 8.5 Digit Multimeter
	1 Hz to 2 kHz	0.007 6 % of reading + 50 μV	
	(2 to 10) kHz	0.012 % of reading + 50 μV	
	(10 to 30) kHz	0.023 % of reading + 0.1 mV	
	(30 to 100) kHz	0.053 % of reading + 0.5 mV	
	(100 to 300) kHz	0.21 % of reading + 3.1 mV	
	300 kHz to 1 MHz	1 % of reading + 10 mV	
	(1 to 2) MHz	1.5 % of reading + 50 mV	
	(2 to 4) MHz	4 % of reading + 0.1 V	
	(4 to 8) MHz	8.2 % of reading + 0.1 V	
	(8 to 10) MHz	15 % of reading + 0.1 V	
	(10 to 100) V		
	1 Hz to 2 kHz	0.009 % of reading + 0.5 mV	
	(2 to 10) kHz	0.011 % of reading + 0.5 mV	
	(10 to 30) kHz	0.023 % of reading + 1 mV	
(30 to 100) kHz	0.059 % of reading + 5 mV		
(100 to 300) kHz	0.37 % of reading + 47 mV		
300 kHz to 1 MHz	1 % of reading + 0.5 V		
(100 to 1 050) V			
1 Hz to 2 kHz	0.011 % of reading + 25 mV		
(2 to 10) kHz	0.011 % of reading + 25 mV		
(10 to 30) kHz	0.023 % of reading + 25 mV		
(30 to 100) kHz	0.059 % of reading + 0.1 V		
DC Current – Source ¹	Up to 120 μA	0.009 8 % of reading + 4.7 nA	Fluke 5560A Multiproduct Calibrator
	(0.12 to 1.2) mA	0.007 8 % of reading + 12 nA	
	(1.2 to 12) mA	0.007 8 % of reading + 63 nA	
	(12 to 120) mA	0.007 8 % of reading + 0.63 μA	
	(0.12 to 1.2) A	0.013 % of reading + 7.8 μA	
	(1.2 to 3.1) A	0.024 % of reading + 0.12 mA	
	(3.1 to 12) A	0.024 % of reading + 0.2 mA	
(12 to 30.2) A	0.078 % of reading + 0.4 mA		
DC Current – Measure ¹	Up to 20 μA	0.002 9 % of reading + 0.4 nA	Fluke 8588A 8.5 Digit Multimeter
	(20 to 200) μA	0.001 % of reading + 0.39 nA	
	(0.2 to 2) mA	0.001 % of reading + 3.9 nA	
	(2 to 20) mA	0.001 5 % of reading + 39 nA	
	(20 to 200) mA	0.005 8 % of reading + 1 μA	
	(0.2 to 2) A	0.013 % of reading + 0.1 mA	
	(2 to 20) A	0.023 % of reading + 0.4 mA	
	(20 to 30) A	0.055 % of reading + 4.4 mA	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Clamp-on Ammeter (Non-Toroidal Type) Hall Effect Sensor ¹	(20 to 60) A (60 to 155) A (155 to 600) A (600 to 1 500) A	0.59 % of reading + 3.9 mA 0.59 % of reading + 6 mA 0.59 % of reading + 9.5 mA 0.6 % of reading + 20 mA	Fluke 5560A Multiproduct Calibrator, Fluke 55XXA/COIL 50 50-turn Coil
DC Voltage – Source ¹	Up to 120 mV (0.12 to 1.2) V (1.2 to 12) V (12 to 120) V (120 to 1 020) V	0.000 93 % of reading + 0.62 μV 0.000 64 % of reading + 0.78 μV 0.000 62 % of reading + 7.8 μV 0.000 85 % of reading + 78 μV 0.000 86 % of reading + 0.78 mV	Fluke 5560A Multiproduct Calibrator
DC Voltage – Measure ¹	Up to 200 mV (0.2 to 2) V (2 to 20) V (20 to 200) V (200 to 1050) V	0.000 77 % of reading + 0.2 μV 0.000 29 % of reading + 0.3 μV 0.000 29 % of reading + 0.47 μV 0.000 43 % of reading + 30 μV 0.000 44 % of reading + 0.5 mV	Fluke 8588A 8.5 Digit Multimeter
Resistance – Source ¹ (Simulation)	Up to 12 Ω (12 to 120) Ω (0.12 to 1.2) kΩ (1.2 to 12) kΩ (12 to 120) kΩ (0.12 to 1.2) MΩ (1.2 to 12) MΩ (12 to 120) MΩ (0.12 to 1.2) GΩ	0.002 % of reading + 0.78 mΩ 0.002 % of reading + 0.78 mΩ 0.002 % of reading + 1.6 mΩ 0.002 % of reading + 16 mΩ 0.002 % of reading + 0.16 Ω 0.002 % of reading + 1.6 Ω 0.002 8 % of reading + 24 Ω 0.034 % of reading + 2 kΩ 0.32 % of reading + 78 kΩ	Fluke 5560A Multiproduct Calibrator (4-wire Mode)
Resistance – Measure ¹	Up to 2 Ω (2 to 20) Ω (20 to 200) Ω (0.2 to 2) kΩ (2 to 20) kΩ (20 to 200) kΩ (0.2 to 2) MΩ (2 to 20) MΩ (20 to 200) MΩ (0.2 to 2) GΩ	0.001 6 % of reading + 4 μΩ 0.001 % of reading + 14 μΩ 0.000 92 % of reading + 47 μΩ 0.000 91 % of reading + 0.47 mΩ 0.000 93 % of reading + 4.7 mΩ 0.000 93 % of reading + 47 mΩ 0.001 1 % of reading + 1 Ω 0.001 9 % of reading + 0.1 kΩ 0.012 % of reading + 10 kΩ 0.13 % of reading + 1 MΩ	Fluke 8588A 8.5 Digit Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Low Current Resistance – Measure ¹	Up to 2 Ω (2 to 20) Ω (20 to 200) Ω (0.2 to 2) kΩ (2 to 20) kΩ (20 to 200) kΩ (0.2 to 2) MΩ (2 to 20) MΩ (20 to 200) MΩ (0.2 to 2) GΩ	0.001 7 % of reading + 4 μΩ 0.001 % of reading + 14 μΩ 0.001 7 % of reading + 0.2 mΩ 0.001 8 % of reading + 2 mΩ 0.002 2 % of reading + 20 mΩ 0.002 2 % of reading + 62 mΩ 0.002 6 % of reading + 1 Ω 0.038 % if reading + 0.3 kΩ 0.013 % of reading + 10 kΩ 0.13 % of reading + 1 MΩ	Fluke 8588A 8.5 Digit Multimeter
High Voltage Resistance – Measure ¹	(2 to 20) MΩ (20 to 200) MΩ (0.2 to 2) GΩ (2 to 20) GΩ	0.001 7 % of reading + 10 Ω 0.006 8 % of reading + 0.1 kΩ 0.023 % of reading + 0.1 MΩ 0.13 % of reading + 10 MΩ	Fluke 8588A 8.5 Digit Multimeter
Capacitance – Source ¹ (Simulation)	Up to 1.2 nF (20 to 40) Hz (40 to 100) Hz 100 Hz to 10 kHz (10 to 12) kHz (12 to 14) kHz (1.2 to 3) nF (10 to 150) Hz 150 Hz to 5 kHz (5 to 6) kHz (6 to 8) kHz (3 to 12) nF 10 Hz to 5 kHz (5 to 6) kHz (6 to 8) kHz (12 to 30) nF (20 to 200) Hz 200 Hz to 1.3 kHz (1.3 to 2.7) kHz (2.7 to 3.7) kHz (30 to 120) nF 10 Hz to 1.3 kHz (1.3 to 2.7) kHz (2.7 to 3.7) kHz (0.12 to 1.2) μF (2 to 310) Hz (310 to 800) Hz 800 Hz to 1.1 kHz	0.87 % of reading + 1.6 pF 0.49 % of reading + 1.6 pF 0.094 % of reading + 1.6 pF 0.48 % of reading + 1.6 pF 0.87 % of reading + 1.6 pF 0.49 % of reading + 3.9 pF 0.09 % of reading + 3.9 pF 0.48 % of reading + 3.9 pF 0.87 % of reading + 3.9 pF 0.09 % of reading + 3.9 pF 0.48 % of reading + 3.9 pF 0.87 % of reading + 3.9 pF 0.49 % of reading + 24 pF 0.1 % of reading + 24 pF 0.49 % of reading + 24 pF 0.88 % of reading + 24 pF 0.1 % of reading + 24 pF 0.49 % of reading + 24 pF 0.88 % of reading + 24 pF 0.1 % of reading + 0.24 nF 0.49 % of reading + 0.24 nF 0.88 % of reading + 0.24 nF	Fluke 5560A Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment			
Capacitance – Source ¹ (Simulation)	(1.2 to 12) μ F 500 mHz to 110 Hz (110 to 250) Hz (250 to 350) Hz	0.1 % of reading + 2.4 nF 0.49 % of reading + 2.4 nF 0.88 % of reading + 2.4 nF	Fluke 5560A Multiproduct Calibrator			
	(12 to 120) μ F (100 to 500) mHz 500 mHz to 40 Hz (40 to 80) Hz (80 to 110) Hz	0.5 % of reading + 20 nF 0.12 % of reading + 20 nF 0.5 % of reading + 20 nF 0.89 % of reading + 20 nF				
	(0.12 to 1.2) mF 100 mHz to 11 Hz (11 to 18) Hz (18 to 25) Hz	0.2 % of reading + 0.2 μ F 0.58 % of reading + 0.2 μ F 1 % of reading + 0.2 μ F				
	(1.2 to 12) mF 30 mHz to 4 Hz (4 to 6) Hz (6 to 8) Hz	0.19 % of reading + 2.4 μ F 0.58 % of reading + 2.4 μ F 1 % of reading + 2.4 μ F				
	(12 to 120) mF 10 mHz to 1.3 Hz (1.3 to 1.7) Hz (1.7 to 2.5) Hz	0.39 % of reading + 24 μ F 0.78 % of reading + 24 μ F 1.2 % of reading + 24 μ F				
	Capacitance – Measure ¹	Up to 2 nF (2 to 20) nF (20 to 200) nF (0.2 to 2) μ F (2 to 20) μ F (20 to 200) μ F (0.2 to 2) mF (2 to 20) mF (20 to 200) mF		0.19 % of reading + 1 pF 0.081 % of reading + 2 pF 0.049 % of reading + 10 pF 0.041 % of reading + 0.1 nF 0.042 % of reading + 1 nF 0.061 % of reading + 10 nF 0.061 % of reading + 0.1 μ F 0.071 % of reading + 1 μ F 0.072 % of reading + 10 μ F	Fluke 8588A 8.5 Digit Multimeter	
		Inductance – Source ¹ (Simulation)		Up to 120 μ H (490 to 550) Hz 550 Hz to 1 kHz 1 kHz (1 to 13) kHz (13 to 17) kHz	0.93 % of reading + 0.16 μ H 0.35 % of reading + 0.16 μ H 0.16 % of reading + 0.16 μ H 0.35 % of reading + 0.16 μ H 0.93 % of reading + 0.16 μ H	Fluke 5560A Multiproduct Calibrator
				(0.12 to 1.2) mH (260 to 330) Hz 330 Hz to 1 kHz 1 kHz (1 to 1.6) kHz (1.6 to 2.5) kHz	0.87 % of reading + 0.78 μ H 0.29 % of reading + 0.78 μ H 0.094 % of reading + 0.78 μ H 0.29 % of reading + 0.78 μ H 0.87 % of reading + 0.78 μ H	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Inductance – Source ¹ (Simulation)	(1.2 to 3.3) mH		Fluke 5560A Multiproduct Calibrator
	500 mHz to 110 Hz	0.3 % of reading + 7.8 μH	
	110 Hz	0.094 % of reading + 7.8 μH	
	(110 to 800) Hz	0.29 % of reading + 7.8 μH	
	(800 to 980) Hz	0.87 % of reading + 7.8 μH	
	(3.3 to 12) mH		
	500 mHz to 110 Hz	0.29 % of reading + 7.8 μH	
	110 Hz	0.093 % of reading + 7.8 μH	
	110 to 1 kHz	0.29 % of reading + 7.8 μH	
	(1 to 1.4) kHz	0.87 % of reading + 7.8 μH	
	(12 to 83) mH		
	100 mHz to 100 Hz	0.29 % of reading + 78 μH	
	100 Hz	0.093 % of reading + 78 μH	
	(100 to 180) Hz	0.29 % of reading + 78 μH	
	(180 to 230) Hz	0.87 % of reading + 78 μH	
	(83 to 120) mH		
	100 mHz to 100 Hz	0.29 % of reading + 78 μH	
	100 Hz	0.093 % of reading + 78 μH	
	(100 to 320) Hz	0.3 % of reading + 78 μH	
	320 Hz to 1 kHz	0.87 % of reading + 78 μH	
	(120 to 650) mH		
	50 mHz to 10 Hz	0.32 % of reading + 0.78 mH	
	10 Hz	0.14 % of reading + 0.78 mH	
	(10 to 30) Hz	0.32 % of reading + 0.78 mH	
	(30 to 55) Hz	0.9 % of reading + 0.78 mH	
	(0.65 to 1.2) H		
	50 mHz to 10 Hz	0.32 % of reading + 0.78 mH	
	10 Hz	0.14 % of reading + 0.78 mH	
	(10 to 100) Hz	0.32 % of reading + 0.78 mH	
	(100 to 170) Hz	0.9 % of reading + 0.78 mH	
(1.2 to 5.5) H			
10 mHz to 3 Hz	0.36 % of reading + 7.8 mH		
3 Hz	0.17 % of reading + 7.8 mH		
(3 to 8) Hz	0.36 % of reading + 7.8 mH		
(8 to 16) Hz	0.93 % of reading + 7.8 mH		
(5.5 to 12) H			
10 mHz to 3 Hz	0.36 % of reading + 7.8 mH		
3 Hz	0.17 % of reading + 7.8 mH		
(3 to 19) Hz	0.36 % of reading + 7.8 mH		
(19 to 37) Hz	0.93 % of reading + 7.8 mH		

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Inductance – Source ¹ (Simulation)	(12 to 30) H 5 mHz to 2 Hz 2 Hz (2 to 4) Hz (4 to 9) Hz	0.39 % of reading + 78 mH 0.21 % of reading + 78 mH 0.39 % of reading + 78 mH 1 % of reading + 78 mH	Fluke 5560A Multiproduct Calibrator
	(30 to 120) H 5 mHz to 2 Hz 2 Hz (2 to 7) Hz (7 to 14) Hz	0.39 % of reading + 78 mH 0.21 % of reading + 78 mH 0.39 % of reading + 78 mH 1 % of reading + 78 mH	
Phase – Source ¹	(0 to 360) °		Fluke 5560A Multiproduct Calibrator
	(3 to 65) Hz	0.08 °	
	(65 to 500) Hz	0.19 °	
	500 Hz to 1 kHz	0.39 °	
	(1 to 5) kHz	1.9 °	
Electrical Simulation of RTD Indicating Devices – Source ¹	Pt 385, 100 Ω		Fluke 5560A Multiproduct Calibrator
	(-200 to -80) °C	0.039 °C	
	(-80 to 0) °C	0.039 °C	
	(0 to 100) °C	0.054 °C	
	(100 to 300) °C	0.07 °C	
	(300 to 400) °C	0.078 °C	
	(400 to 630) °C	0.093 °C	
	(630 to 800) °C	0.18 °C	
	Pt 385, 200 Ω		
	(-200 to -80) °C	0.031 °C	
	(-80 to 0) °C	0.031 °C	
	(0 to 100) °C	0.031 °C	
	(100 to 260) °C	0.039 °C	
	(260 to 300) °C	0.093 °C	
(300 to 400) °C	0.1 °C		
(400 to 600) °C	0.11 °C		
(600 to 630) °C	0.12 °C		

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of RTD Indicating Devices – Source ¹	Pt 385, 500 Ω		Fluke 5560A Multiproduct Calibrator
	(-200 to -80) °C	0.031 °C	
	(-80 to 0) °C	0.039 °C	
	(0 to 100) °C	0.039 °C	
	(100 to 260) °C	0.047 °C	
	(260 to 300) °C	0.062 °C	
	(300 to 400) °C	0.062 °C	
	(400 to 600) °C	0.07 °C	
	(600 to 630) °C	0.085 °C	
	Pt 385, 1 000 Ω		
	(-200 to -80) °C	0.023 °C	
	(-80 to 0) °C	0.023 °C	
	(0 to 100) °C	0.031 °C	
	(100 to 260) °C	0.039 °C	
	(260 to 300) °C	0.047 °C	
	(300 to 400) °C	0.054 °C	
	(400 to 600) °C	0.054 °C	
	(600 to 630) °C	0.18 °C	
	Pt 3916, 100 Ω		
	(-200 to -190) °C	0.19 °C	
	(-190 to -80) °C	0.031 °C	
	(-80 to 0) °C	0.039 °C	
	(0 to 100) °C	0.047 °C	
	(100 to 260) °C	0.054 °C	
	(260 to 300) °C	0.062 °C	
	(300 to 400) °C	0.07 °C	
	(400 to 600) °C	0.078 °C	
	(600 to 630) °C	0.18 °C	
	Pt 3926, 100 Ω		
	(-200 to -80) °C	0.039 °C	
(-80 to 0) °C	0.039 °C		
(0 to 100) °C	0.054 °C		
(100 to 300) °C	0.07 °C		
(300 to 400) °C	0.078 °C		
(400 to 630) °C	0.093 °C		
Ni 672, 120 Ω			
(-80 to 0) °C	0.06 °C		
(0 to 100) °C	0.06 °C		
(100 to 260) °C	0.11 °C		

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of RTD Indicating Devices – Source ¹	Cu 427, 10 Ω (-80 to 260) °C	0.23 °C	Fluke 5560A Multiproduct Calibrator
	Cu 428, 50 Ω (-180 to 200) °C	0.31 °C	
	Cu 428, 100 Ω (-180 to 40) °C	0.31 °C	
	(40 to 200) °C	0.5 °C	
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure ¹	Type B (600 to 800) °C	0.34 °C	Fluke 5560A Multiproduct Calibrator
	(800 to 1 000) °C	0.26 °C	
	(1 000 to 1 550) °C	0.23 °C	
	(1 550 to 1 820) °C	0.26 °C	
	Type C (0 to 150) °C	0.19 °C	
	(150 to 650) °C	0.16 °C	
	(650 to 1 000) °C	0.2 °C	
	(1 000 to 1 800) °C	0.35 °C	
	(1 800 to 2 315) °C	0.61 °C	
	Type D (0 to 150) °C	0.19 °C	
	(150 to 650) °C	0.16 °C	
	(650 to 1 000) °C	0.2 °C	
	(1 000 to 1 800) °C	0.34 °C	
	(1 800 to 2 315) °C	0.61 °C	
	Type E (-250 to -150) °C	0.31 °C	
	(-150 to -25) °C	0.11 °C	
	(-25 to 350) °C	0.09 °C	
	(350 to 650) °C	0.12 °C	
	(650 to 1 000) °C	0.16 °C	
	Type G (0 to 150) °C	0.39 °C	
	(150 to 650) °C	0.26 °C	
	(650 to 1 000) °C	0.2 °C	
	(1 000 to 1 800) °C	0.33 °C	
	(1 800 to 2 315) °C	0.6 °C	
Type J (-210 to -100) °C	0.19 °C		
(-100 to -30) °C	0.1 °C		
(-30 to 150) °C	0.09 °C		
(150 to 760) °C	0.11 °C		
(760 to 1 200) °C	0.16 °C		

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure ¹	Type K		Fluke 5560A Multiproduct Calibrator
	(-200 to -100) °C	0.22 °C	
	(-100 to -25) °C	0.1 °C	
	(-25 to 120) °C	0.09 °C	
	(120 to 1 000) °C	0.16 °C	
	(1 000 to 1 372) °C	0.27 °C	
	Type L		
	(-200 to -100) °C	0.24 °C	
	(-100 to 800) °C	0.16 °C	
	(800 to 900) °C	0.09 °C	
	Type N		
	(-200 to -100) °C	0.26 °C	
	(-100 to -25) °C	0.12 °C	
	(-25 to 120) °C	0.09 °C	
	(120 to 410) °C	0.09 °C	
	(410 to 1 300) °C	0.16 °C	
	Type R		
	(0 to 250) °C	0.4 °C	
	(250 to 400) °C	0.23 °C	
	(400 to 1 000) °C	0.21 °C	
	(1 000 to 1 767) °C	0.26 °C	
	Type S		
	(0 to 250) °C	0.33 °C	
	(250 to 400) °C	0.24 °C	
(400 to 1 000) °C	0.25 °C		
(1 000 to 1 767) °C	0.32 °C		
Type T			
(-250 to -150) °C	0.47 °C		
(-150 to 0) °C	0.16 °C		
(0 to 120) °C	0.1 °C		
(120 to 400) °C	0.09 °C		
Type U			
(-200 to 0) °C	0.31 °C		
(0 to 600) °C	0.09 °C		
Type BP			
(0 to 1 000) °C	0.31 °C		
(1 000 to 2 000) °C	0.47 °C		
(2 000 to 2 500) °C	0.62 °C		
Type XK			
(-200 to 300) °C	0.16 °C		
(300 to 800) °C	0.23 °C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure ¹	Type B		Ectron 1140A Thermocouple Calibrator/Simulator
	(250 to 350) °C	1.2 °C	
	(350 to 445) °C	0.9 °C	
	(445 to 580) °C	0.71 °C	
	(580 to 750) °C	0.55 °C	
	(750 to 1 000) °C	0.45 °C	
	(1 000 to 1 820) °C	0.35 °C	
	Type C		
	(0 to 250) °C	0.24 °C	
	(250 to 1 000) °C	0.19 °C	
	(1 000 to 1 500) °C	0.21 °C	
	(1 500 to 1 800) °C	0.24 °C	
	(1 800 to 2 000) °C	0.27 °C	
	(2 000 to 2 250) °C	0.33 °C	
	(2 250 to 2 315) °C	0.37 °C	
	Type D		
	(0 to 100) °C	0.34 °C	
	(100 to 300) °C	0.28 °C	
	(300 to 1 400) °C	0.18 °C	
	(1 400 to 1 650) °C	0.19 °C	
	(1 650 to 1 930) °C	0.23 °C	
	(1 930 to 2 100) °C	0.28 °C	
	(2 100 to 2 200) °C	0.3 °C	
	(2 200 to 2 320) °C	0.31 °C	
	Type E		
	(-270 to -245) °C	1.6 °C	
	(-245 to -195) °C	0.24 °C	
	(-195 to -155) °C	0.12 °C	
	(-155 to -90) °C	0.09 °C	
	(-90 to 0) °C	0.08 °C	
(0 to 15) °C	0.08 °C		
(15 to 890) °C	0.06 °C		
(890 to 1 000) °C	0.07 °C		
Type G			
(0 to 100) °C	1.6 °C		
(100 to 300) °C	0.5 °C		
(300 to 600) °C	0.35 °C		
(600 to 1 760) °C	0.18 °C		
(1 760 to 2 030) °C	0.2 °C		
(2 030 to 2 200) °C	0.25 °C		
(2 200 to 2 315) °C	0.27 °C		

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure ¹	Type J		Ectron 1140A Thermocouple Calibrator/Simulator
	(-210 to -180) °C	0.15 °C	
	(-180 to -120) °C	0.12 °C	
	(-120 to -50) °C	0.09 °C	
	(-50 to 990) °C	0.08 °C	
	(990 to 1 200) °C	0.09 °C	
	Type K		
	(-270 to -255) °C	2.5 °C	
	(-255 to -195) °C	0.85 °C	
	(-195 to -115) °C	0.16 °C	
	(-115 to -55) °C	0.12 °C	
	(-55 to 1 000) °C	0.09 °C	
	(1 000 to 1 372) °C	0.1 °C	
	Type N		
	(-270 to -260) °C	5.4 °C	
	(-260 to -200) °C	1.5 °C	
	(-200 to -140) °C	0.29 °C	
	(-140 to -70) °C	0.18 °C	
	(-70 to 25) °C	0.14 °C	
	(25 to 160) °C	0.12 °C	
	(160 to 1 300) °C	0.11 °C	
	Type PLII		
	(0 to 100) °C	0.1 °C	
	(100 to 925) °C	0.08 °C	
	(925 to 1 200) °C	0.1 °C	
	(1 200 to 1 395) °C	0.11 °C	
	Type R		
	(-50 to -30) °C	0.8 °C	
(-30 to 45) °C	0.69 °C		
(45 to 160) °C	0.49 °C		
(160 to 380) °C	0.35 °C		
(380 to 775) °C	0.3 °C		
(775 to 1 768) °C	0.26 °C		
Type S			
(-50 to -30) °C	0.76 °C		
(-30 to 45) °C	0.68 °C		
(45 to 105) °C	0.49 °C		
(105 to 310) °C	0.41 °C		
(310 to 615) °C	0.35 °C		
(615 to 1 768) °C	0.31 °C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure ¹	Type T (-270 to -255) °C (-255 to -240) °C (-240 to -210) °C (-210 to -150) °C (-150 to -40) °C (-40 to 100) °C (100 to 400) °C	1.9 °C 0.6 °C 0.36 °C 0.22 °C 0.15 °C 0.09 °C 0.08 °C	Ectron 1140A Thermocouple Calibrator/Simulator
Oscilloscopes ^{1,4} Amplitude – DC into 50 Ω load into 1 MΩ load Amplitude – Square Wave into 50 Ω load into 1 MΩ load Time Markers into 50 Ω load Rise Time into 50 Ω load Rate: 1 kHz to 2 MHz Rate: (2 to 10) MHz	(-6.6 to 6.6) V (-130 to 130) V 1 mVp-p to 6.6 Vp-p 10 Hz to 10 kHz 1 mVp-p to 130 Vp-p 10 Hz to 1 kHz (1 to 10) kHz 1 ns to 20 ms 50 ms 100 ms 200 ms 500 ms 1 s 2 s 5 s 5 mVp-p to 2.5 Vp-p (200 to 300) ps (250 to 350) ps	0.22 % of reading + 31 μV 0.12 % of reading + 31 μV 0.22 % of reading + 31 μV 0.14 % of reading + 31 μV 0.22 % of reading + 31 μV 0.000 22 % of reading 0.005 9 % of reading 0.009 8 % of reading 0.018 % of reading 0.041 % of reading 0.080 % of reading 0.16 % of reading 0.39 % of reading 50 ps 50 ps	Fluke 5560A/1G Multiproduct Calibrator with 1.1 GHz Scope Option



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes ^{1,4} Level Sine Wave into 50 Ω load	5 mVp-p to 5.5 Vp-p 50 kHz 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz 5 mVp-p to 3.5 Vp-p 600 MHz to 1.1 GHz	1.8 % of reading + 0.23 mV 2.8 % of reading + 0.23 mV 3.2 % of reading + 0.23 mV 4 % of reading + 0.23 mV 5.5 % of reading + 0.23 mV	Fluke 5560A/1G Multiproduct Calibrator with 1.1 GHz Scope Option
Bandwidth Flatness (50 kHz Reference) into 50 Ω load	5 mVp-p to 5.5 Vp-p 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz 5 mVp-p to 3.5 Vp-p 600 MHz to 1.1 GHz	1.4 % of reading + 78 μV 1.8 % of reading + 78 μV 3.2 % of reading + 78 μV 4 % of reading + 78 μV	
Input Impedance Measure	(40 to 60) Ω (0.5 to 1.5) MΩ	0.082 % of reading 0.081 % of reading	
Input Capacitance Measure	(5 to 50) pF	3.9 % of reading + 0.39 pF	
Wave Generator (Sine, Square, Triangle) Amplitude into 50 Ω load	1.8 mVp-p to 2.5 Vp-p 10 Hz to 10 kHz	2.3 % of reading + 78 μV	
into 1 MΩ load	1.8 mVp-p to 55 Vp-p 10 Hz to 10 kHz	2.3 % of reading + 78 μV	
Frequency	10 Hz to 10 kHz	0.0019 % of reading + 12 mHz	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Power – Source ^{1,5} (10 to 65) Hz Power Factor = 1	(3.3 to 9) mA (0.11 to 3) mW 3 mW to 9 W (9 to 33) mA (0.3 to 10) mW 10 mW to 33 W (33 to 90) mA (1 to 30) mW 30 mW to 90 W (90 to 330) mA (3 to 100) mW 100 mW to 300 W (0.33 to 0.9) A (11 to 300) mW 300 mW to 900 W (0.9 to 2.2) A 30 mW to 0.72 W 720 mW to 2 kW (2.2 to 4.5) A 80 mW to 1.4 W 1.4 W to 4.5 kW (4.5 to 20.5) A (0.15 to 6.7) W 6.7 W to 20 kW	0.13 % of reading 0.077 % of reading 0.089 % of reading 0.077 % of reading 0.071 % of reading 0.057 % of reading 0.089 % of reading 0.078 % of reading 0.071 % of reading 0.081 % of reading 0.089 % of reading 0.079 % of reading 0.088 % of reading 0.18 % of reading 0.17 % of reading 0.17 % of reading	Fluke 5522A Multiproduct Calibrator
DC Power – Source ^{1,5}	(0.33 to 330) mA 11 μW to 1.1 mW (1.1 to 110) mW (0.11 to 110) W (110 to 330) W (0.33 to 3) A 11 μW to 110 mW (0.11 to 990) W (0.99 to 3) kW (3 to 20.5) A (99 to 990) mW 0.99 W to 6.8 kW (6.8 to 20.5) kW	0.024 % of reading 0.027 % of reading 0.024 % of reading 0.018 % of reading 0.044 % of reading 0.053 % of reading 0.009 6 % of reading 0.088 % of reading 0.07 % of reading 0.04 % of reading	Fluke 5522A Multiproduct Calibrator

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Mass Determination ^{2,3} (OIML Class E2, F1, F2)	1 mg	2 µg	Modified Double Substitution; Electronic Mass Comparators, Master Weight Sets
	2 mg	2 µg	
	5 mg	2 µg	
	10 mg	2.7 µg	
	20 mg	3 µg	
	50 mg	4 µg	
	100 mg	5.3 µg	
	200 mg	6.7 µg	
	500 mg	8.3 µg	
	1 g	10 µg	
	2 g	13 µg	
	5 g	17 µg	
	10 g	20 µg	
	20 g	27 µg	
	50 g	33 µg	
	100 g	53 µg	
	200 g	0.1 mg	
500 g	0.27 mg		
1 kg	0.53 mg		
2 kg	1 mg		
5 kg	2.7 mg		
10 kg	5.3 mg		
Mass Determination ^{2,3} (OIML Class F1, F2)	20 kg	33 mg	Modified Single Substitution; Electronic Mass Comparator, Master Weight
Mass Determination ^{2,3} (OIML Class M1, M2, M3)	1 mg	2 µg	Modified Single Substitution; Electronic Mass Comparators, Master Weight Set
	2 mg	2 µg	
	5 mg	2 µg	
	10 mg	2.7 µg	
	20 mg	3 µg	
	50 mg	4 µg	
	100 mg	5.3 µg	
	200 mg	6.7 µg	
	500 mg	8.3 µg	



ANSI National Accreditation Board

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Mass Determination ^{2,3} (OIML Class M1, M2, M3)	1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg 25 kg	10 µg 13 µg 17 µg 20 µg 27 µg 33 µg 1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 0.1 g 0.2 g 0.25 g	Modified Single Substitution; Electronic Mass Comparators, Master Weight Set
Force Gages (Tension and Compression)	(10 to 300) lbf (300 to 2 000) lbf	0.024 % of reading + 0.041 lbf 0.024 % of reading + 0.33 lbf	Comparison to Morehouse Force Calibration System


Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Source ¹	10 mHz to 2 MHz	0.000 19 % of reading	Fluke 5560A Multiproduct Calibrator
Frequency – Measure ¹	10 Hz to 100 MHz	0.008 % of reading	Fluke 8588A 8.5 Digit Multimeter
AC Duty Cycle – Source ¹ Square Wave: < 3.3 V _{p-p} Freq.: 10 mHz to 100 kHz	(10 to 49) % Duty Cycle 10 µs to 100 s 50 % Duty Cycle 10 µs to 100 s (51 to 90) % Duty Cycle 10 µs to 100 s	0.039 % of reading + 78 ns 0.016 % of reading + 78 ns 0.039 % of reading + 78 ns	Fluke 5560A Multiproduct Calibrator
Stopwatches. Timers ¹	Up to 19.99 s/d	58 ms/d	Direct Measurement based on NIST SP 960-12; Helmut Klein/Vibrograf TM-4500 Timometer

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. Laboratory maintains the ability to generate quantities between the reported uncertainties. This reported uncertainty may be larger than the cardinal points listed within this parameter.
3. Newton and Pound weights are mathematically converted to equivalent kilograms.
4. The stated uncertainty is the laboratory's ability to source a fast rise pulse that is approximately 250 ps. In the typical application of measuring rise time of an oscilloscope, this value is one of the contributing factors, but other factors are derived from the DUT.
5. The uncertainties shown are for the most favorable conditions. There is an increase in uncertainty that corresponds to the laboratory's AC voltage and current uncertainties at different frequencies other than the ones shown. Power factors (PF) other than the one shown contribute to the power uncertainty. PF is related to the cosine of phase. Therefore, uncertainties track the laboratory's phase uncertainty closely at PF near one but are magnified heavily as PF approaches zero. The lab may also report reactive power, apparent power, and power factor under this accreditation. If needed, contact the laboratory for more information regarding uncertainties at frequency and power factor combinations other than the ones shown.
6. Unless otherwise specified in the Reference Standard, Method, and/or Equipment column, the calibration procedure or method used during the calibration of these parameters were developed internally and were validated according to ISO/IEC 17025:2017 by the CAB.
7. The legal entity for this client is Transcat, Inc.
8. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2489.29.



Jason Stine, Vice President

